

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 16 2010

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130112</u> ✓	
Name <u>Bruce Smith Jr</u>	ID# <u>3522</u> Date <u>6-8-10</u>
A Agency <u>Wrangell Police Department</u> Phone # <u>(907) 874-3304</u>	
Instrument Location <u>Wrangell Police Department, 431 Zimovia Highway</u> ✓	
B Alco S/N <u>77199</u> ✓ Target Value <u>.083</u> ✓ High Pressure <u>800</u>	
Alco Test Values <u>.088</u> ✓ <u>.088</u> ✓ 1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u> ✓ <u>COB</u> <u>6/18/10</u>	
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 24<sup>th</sup> day of June, 2010.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland,  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130112 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130112 ✓

JUNE 08, 2010 ✓

OPERATOR'S NAME:

SMITH JR, BRUCE W

OPERATOR'S NUMBER: 3522

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: 0123456

DEPT/AGENCY: MRG1

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .083 ✓

ALCO S/N: 77199 ✓

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 30.30 in ✓  
ALCO TARGET .084 19:01  
BLANK TEST .000 19:02  
INTERNAL STANDARD VERIFIED 19:02  
ALCO TV 30.30 in .083 19:02  
BLANK TEST .000 ✓ 19:03  
SUBJECT SAMPLE .000 19:03  
BLANK TEST .000 ✓ 19:04  
ALCO TV 30.30 in .083 ✓ 19:04  
BLANK TEST .000 19:05

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130112 ✓

JUNE 08, 2010 ✓

TIME 18:59

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 48c  
BREATH TUBE: 42c  
BAROMETER: 30.30 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!\"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~  
pqrstuvwxyzi!~